

**KERALA AGRICULTURAL UNIVERSITY  
KELAPPAJI COLLEGE OF AGRI ENGG & FOOD TECHNOLOGY**

Tavanur-679573, Malappuram (Dt), Kerala  
Email: deanengg@kau.in, kcaet@kau.in; Ph: 0494 2686214

No. KCAET/1339/2025-B3

Dated: 03-03-2025

QUOTATION NOTICE

Sealed competitive quotations are invited for the supply of Water Quality Analyzer for academic and research activities of this college, subject to the terms and conditions stated below.

Sl No	Item	Specification	Qty
1	<b>Water Quality Analyzer</b> (To measure pH, Conductivity/TDS, Salinity, Dissolved Oxygen, Temperature and Turbidity)	Temperature : 0 <sup>0</sup> C to 100 <sup>0</sup> C pH : 0 to 14 Conductivity/TDS : 0 to 100 milli Siemens Salinity : 0 to 0.00004 ppm Dissolved oxygen : 0 to 20 ppm Turbidity : 0 to 100 NTU	1 No.
Warranty: Minimum 2 years			

Terms and Conditions

1. The rates quoted should be inclusive of all taxes and other charges if any and the items shall be delivered and installed at the points specified by the undersigned
2. The rate quoted should be that of reputed firm having quality standards.
3. The quotations should reach this office before 12.00 pm on 12-03-2025 and the quotation will be opened at 12.30 pm on the same day.
4. Late and defective quotations will not be accepted
5. The successful quotationer is liable to submit a security deposit @ 5% of total cost at the time of installation itself without which the settlement of invoice will not be done
6. The warranty /post supply service details etc. should be mentioned clearly.
7. The envelope containing the quotation should be super scribed with **“Quotation for the supply of Water Quality Analyzer”** for academic and research activities of this college as per order No. KCAET/1339/2025-B3” and addressed to the Dean of faculty (Ag. Engg), KCAET, Tavanur, Malappuram (Dt) Pin-679 573.
8. All rules related to Government quotations will be applicable in this case also.

DEAN OF FACULTY (AGRICULTURAL ENGINEERING)

Signed by  
Jayan P.R  
Jayan P.R

Date: 03-03-2025 12:56:

To : KAU Website, KCAET Website  
Notice Board